




measurement



calibration



as-delivered condition



subject to technical alterations

type: aerosol  
range: 2,5 - 8 mm  
range of indication:  
digital step: 0,01 mm  
scale interval:  
measuring force: 0,7 - 1,3 N  
weight: ca.250 g

acceptable deviations  
VDI/VDE/DGQ 2618 Bl. 12.1/13.1)  
max. permissible  
errors "G": 0,01 mm  
repeatability "r": 0,01 mm  
reference temperature: 20° C  
operation temperature: 10° C - 30° C  
storage temperature: -10° C - 50° C

scale: 1:1  
drawing-nr.: IDW-AE1180-DB\_e  
date of issue: 25.11.2010  
name: J.Heuser  
revision status: 2310 (CT)  
revision date: 06.03.2024